

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 4254
CALIBRATION DATE: 01-Dec-16

SBE 37 RTC CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

rtca0 = 9.999813e-001

rtca1 = 1.646957e-006

rtca2 = -3.302787e-008

BATH TEMPERATURE (° C)	RTC FREQUENCY (Hz)	COMPUTED FREQUENCY (Hz)	RESIDUAL (PPM)
1.0000	0.9999830	0.9999829	-0.1
4.5000	0.9999880	0.9999880	0.0
15.0000	0.9999980	0.9999986	0.6
18.5000	1.0000010	1.0000005	-0.5
24.0000	1.0000020	1.0000018	-0.2
28.9999	1.0000010	1.0000013	0.3
32.5000	1.0000000	0.9999999	-0.1

t = temperature (°C)

RTC frequency = $rtca0 + rtca1 * t + rtca2 * t^2$

Residual (PPM) = (Computed RTC frequency - Measured RTC frequency) * 1e6

