

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA
 Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 4303
 CALIBRATION DATE: 30-Jan-08

SBE 37 PRESSURE CALIBRATION DATA
 10153 psia S/N 8843

COEFFICIENTS:

PA0 = 1.089311e+000
 PA1 = 4.892954e-001
 PA2 = -4.885670e-007

PTCA0 = -6.059328e+001
 PTCA1 = 1.907403e-001
 PTCA2 = -4.063664e-003
 PTCB0 = 2.484362e+001
 PTCB1 = -1.475000e-003
 PTCB2 = 0.000000e+000

PRESSURE SPAN CALIBRATION

PRESSURE PSIA	INST OUTPUT	TEMP ITS90	COMPUTED PRESSURE	ERROR %FS
14.56	-30.8	21.6	14.60	0.00
2001.49	4039.5	21.7	2000.52	-0.01
3988.55	8146.2	21.6	3987.73	-0.01
5975.67	12288.7	21.8	5975.61	-0.00
7962.72	16465.8	21.8	7963.02	0.00
9950.61	20677.9	21.8	9949.77	-0.01
7962.81	16467.3	21.8	7963.74	0.01
5975.49	12289.8	21.9	5976.17	0.01
3988.39	8147.7	21.9	3988.52	0.00
2001.34	4040.8	21.8	2001.16	-0.00
14.57	-29.2	22.0	15.38	0.01

THERMAL CORRECTION

TEMP ITS90	INST OUTPUT	TEMP ITS90	SPAN MV
32.50	-30.48	-5.00	24.85
29.00	-30.21	35.00	24.79
24.00	-30.09		
18.50	-30.25		
15.00	-30.44		
4.50	-31.53		
1.00	-32.20		

$$x = \text{pressure output} - \text{PTCA0} - \text{PTCA1} * t - \text{PTCA2} * t^2$$

$$n = x * \text{PTCB0} / (\text{PTCB0} + \text{PTCB1} * t + \text{PTCB2} * t^2)$$

$$\text{pressure (psia)} = \text{PA0} + \text{PA1} * n + \text{PA2} * n^2$$

Date, Avg Delta P %FS

30-Jan-08 0.00

