



SEA-BIRD ELECTRONICS, INC.

13431 NE 20th St. Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

Service

Report

RMA Number

72026

Customer Information:

Company Monterey Bay Aquarium Research Institute **Date** 12/20/2012

Contact Jared Figurski

PO Number TBD

Serial Number 37SMP40327-4303

Model Number SBE 37SMP

Services Requested:

1. Evaluate/Repair Instrumentation.
2. Perform Routine Calibration Service.

Problems Found:

1. The anti-foulant devices appeared "dirty". Being installed at customers request.

Services Performed:

1. Performed initial diagnostic evaluation.
2. Calibrated the pressure sensor.
3. Performed "Post Cruise" calibration of the temperature & conductivity sensors.
4. Installed NEW AF24173 Anti-foulant cylinder(s).
5. Performed complete system check and full diagnostic evaluation.

Special Notes: