

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 4303
CALIBRATION DATE: 07-Dec-12

SBE 37 PRESSURE CALIBRATION DATA
10153 psia S/N 8843

COEFFICIENTS:

PA0 = 1.211849e+000
PA1 = 4.894177e-001
PA2 = -4.885041e-007

PTCA0 = -5.976543e+001
PTCA1 = 2.046657e-001
PTCA2 = -4.188135e-003
PTCB0 = 2.484362e+001
PTCB1 = -1.475000e-003
PTCB2 = 0.000000e+000

PRESSURE SPAN CALIBRATION

PRESSURE PSIA	INST OUTPUT	TEMP ITS90	COMPUTED PRESSURE	ERROR %FS
14.60	-29.7	21.9	14.73	0.00
2000.69	4038.4	22.0	2000.12	-0.01
3987.50	8143.3	22.0	3987.04	-0.00
5974.91	12284.6	22.0	5974.85	-0.00
7961.51	16459.1	22.0	7961.59	0.00
9949.23	20670.4	22.1	9948.61	-0.01
7965.17	16468.3	22.1	7965.99	0.01
5972.62	12281.1	22.0	5973.17	0.01
3989.92	8148.9	22.1	3989.76	-0.00
2003.27	4044.5	22.2	2003.11	-0.00
14.58	-29.1	22.2	15.02	0.00

THERMAL CORRECTION

TEMP ITS90	INST OUTPUT	TEMP ITS90	SPAN MV
32.50	-29.21	-5.00	24.85
29.00	-28.97	35.00	24.79
24.00	-28.88		
18.50	-29.07		
15.00	-29.33		
4.50	-30.52		
1.00	-31.24		

$$x = \text{pressure output} - \text{PTCA0} - \text{PTCA1} * t - \text{PTCA2} * t^2$$

$$n = x * \text{PTCB0} / (\text{PTCB0} + \text{PTCB1} * t + \text{PTCB2} * t^2)$$

$$\text{pressure (psia)} = \text{PA0} + \text{PA1} * n + \text{PA2} * n^2$$

Date, Avg Delta P %FS

07-Dec-12 0.00

