



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 4590
CALIBRATION DATE: 19-Dec-17

SBE 37 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.011278e+000
h = 1.407870e-001
i = -1.117455e-004
j = 3.215842e-005

CPcor = -9.5700e-008
CTcor = 3.2500e-006
WBOTC = -6.8062e-006

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2680.97	0.00000	0.00000
1.0000	34.8750	2.98050	5318.86	2.98051	0.00001
4.5000	34.8552	3.28803	5519.10	3.28802	-0.00001
15.0000	34.8131	4.27126	6114.65	4.27126	-0.00001
18.5000	34.8041	4.61693	6310.39	4.61693	0.00001
24.0000	34.7941	5.17569	6614.24	5.17569	0.00000
29.0000	34.7879	5.69819	6885.91	5.69819	-0.00000
32.5000	34.7844	6.07106	7073.20	6.07106	-0.00000

$f = \text{Instrument Output(Hz)} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$

$t = \text{temperature } (°\text{C}); \quad p = \text{pressure (decibars)}; \quad \delta = \text{CTcor}; \quad \epsilon = \text{CPcor};$

$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$

$\text{Residual (Siemens/meter)} = \text{instrument conductivity} - \text{bath conductivity}$

