



SEA-BIRD ELECTRONICS, INC.

1808 - 136th Place Northeast, Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

Pressure Test Certificate

Customer *Monterey Bay Aquarium Research Institute*

Job Number **49503**

Date 2/25/2008

Technician DG

<u>Serial Number</u>	3922160-0225
-----------------------------	--------------

<u>Low Pressure (PSI)</u>	50	PSI
----------------------------------	----	-----

<u>Time (Minutes)</u>	15	Minutes
------------------------------	----	---------

<u>High Pressure (PSI)</u>	500	PSI
-----------------------------------	-----	-----

<u>Time (Minutes)</u>	30	Minutes
------------------------------	----	---------

Pass ☒

Fail ☐

Comments

Replaced the main piston "O"-Rings

