



SEA-BIRD ELECTRONICS, INC.

1808 - 136th Place Northeast, Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

Pressure Test Certificate

Customer *Monterey Bay Aquarium Research Institute*

Job Number **54727**

Date 6/11/2009

Technician DG

Serial Number 4534338-0087

Low Pressure (PSI) 30 PSI

Time (Minutes) 30 Minutes

High Pressure (PSI) *N/A PSI

Time (Minutes) *N/A Minutes

Pass ☒

Fail ☐

Comments

Replaced the main piston "O"-Rings.

*The unit is not designed for high pressure applications.

